



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 055E

PROCUREMENT OF 17 UNITS CAUTERY MACHINE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Recipient:	
Health Facility in Paranaque City	
From	To
RF Frequency: 450kHz \pm 50kHz (Open Circuit)RF Frequency: 450kHz \pm 50kHz (Open Circuit)	RF Frequency: 450kHz \pm 50kHz (Open Circuit)


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9th day of April, 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson *us*




Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS																								
Item No. 1	CAUTERY MACHINE	Qty./Unit	17 Unit																					
Name of Manufacturer:			Country of Origin																					
Brand:			Model: (if applicable)																					
ABC: 850,000.00																								
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE																					
A. Technical Specifications: <ul style="list-style-type: none"> Primary Power Requirement: 220V, 50/60Hz Functional Characteristics: <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 10px;"> <thead> <tr> <th style="width: 10%;">Mode</th> <th style="width: 10%;">Max Power (W)</th> <th style="width: 10%;">Max Voltage (Vpp)</th> <th style="width: 10%;">Rated Load (Ohms)</th> <th style="width: 10%;">Freq. (kHz)</th> </tr> </thead> <tbody> <tr> <td>HI</td> <td>35</td> <td>8,000</td> <td>1,000</td> <td>24.4</td> </tr> <tr> <td>LO</td> <td>20</td> <td>3,000</td> <td>500</td> <td>32.3</td> </tr> <tr> <td>BI</td> <td>35</td> <td>3,000</td> <td>500</td> <td>32.3</td> </tr> </tbody> </table> Power Accuracy at Rated Load <ul style="list-style-type: none"> - LO and BI: 0 to 10W; Greater than 90mW or 10% of setting - 10 to maximum setting: 10% of setting - HI: Greater of 1W or 10% of setting RF Frequency: 450kHz \pm 50kHz (Open Circuit) Output Wave Shape: Clipped, Damped Sinusoidal Line Regulation: \pm5% at max. power for a \pm10% line voltage Duty Cycle: Intermittent: 30 sec ON, 30 sec OFF Audio: Adjustable from >45dBA to >65dBA at 1 meter 			Mode	Max Power (W)	Max Voltage (Vpp)	Rated Load (Ohms)	Freq. (kHz)	HI	35	8,000	1,000	24.4	LO	20	3,000	500	32.3	BI	35	3,000	500	32.3		
Mode	Max Power (W)	Max Voltage (Vpp)	Rated Load (Ohms)	Freq. (kHz)																				
HI	35	8,000	1,000	24.4																				
LO	20	3,000	500	32.3																				
BI	35	3,000	500	32.3																				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<ul style="list-style-type: none"> • Display Setting <ul style="list-style-type: none"> - Low Output Terminal <ul style="list-style-type: none"> 0-10 range: 0.1-watt increments 10-20 range: 1-watt increments - Bipolar & High Output Terminals <ul style="list-style-type: none"> 0-35 range: 1-watt increments • Environmental <ul style="list-style-type: none"> - Operating Temperature Range: +10°C to +40°C - Relative Humidity: 10% to 95% non-condensing 		
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2; Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3.Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p>		



<p>a) Service manual in English language</p> <p>b) Operation manual in English language</p> <p>6. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit</p> <p>x-</p> <p>Recipient:</p> <p>Health Facility in Paranaque City</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		