



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 130E

PROCUREMENT OF 1 UNIT COLONOSCOPE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Material – fiber Type – flexible Camera – 440, 000 Pixels high resolution Photosensitive CCD Clamp aperture: Φ of 3.7 mm Diameter of distal- end: Φ 13.0mm Total Length: 1, 635mm Depth of view 3-100mm -Must be compatible with the existing equipment of the hospital (PentaX EPK-1000)	
FROM	TO
Submission of Bids (Bid Opening): OCTOBER 19, 2022 AT 9:00 AM	Submission of Bids (Bid Opening): OCTOBER 24, 2022 AT 9:00 AM
• Insertion Tube Diameter: maximum of 11.6 mm	Diameter of insert tube: Φ 12.8mm
• Working Length: 1,700	Working Length: 1,335mm
• Angulation: 180/180	Angulation or Tip deflection: 180°/180° up and down
• Angulation: 160/160	Angulation: 160°/160° right and left
• Angle of view: 140°	Angle of view or view of field: 140°
• Product life: Minimum of 4-5 years	• Product life: Minimum of 4-5 years - removed
REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 – 60 calendar days upon receipt of Notice to Proceed.


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 24th day of October 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17th day of October 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	COLONOSCOPE	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,560,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>A. Technical Specifications:</p> <p>Material – fiber</p> <p>Type – flexible</p> <p>Camera – 440,000 Pixels high resolution Photosensitive CCD</p> <p>Channel Diameter: maximum of 3.8 mm</p> <p>Clamp aperture: Φ of 3.7 mm</p> <p>Diameter of distal- end: Φ13.0mm</p> <p>Diameter of insert tube: Φ12.8mm</p> <p>Working Length: 1,335mm</p> <p>Total Length: 1,635mm</p> <p>Angulation or Tip deflection: 180°/180° up and down</p> <p>Angulation: 160°/160° right and left</p> <p>Angle of view or view of field: 140°</p> <p>Depth of view 3-100mm</p> <p>-Must be compatible with the existing equipment of the hospital (PentaX EPK-1000)</p>			
<p><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 – 60 calendar days upon receipt of Notice to Proceed.</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
3. **Training:** The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.
4. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
 - a) Service manual in English language
 - b) Operation manual in English language
7. With "**DOH-MMCHD HFEP**" (Government Property not for sale) sticker in each unit.

X-X

Recipient: Caloocan City Medical Center – 1 unit

B. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.