



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 133E

#### PROCUREMENT OF 17 UNITS PULSE OXIMETER WITH ADULT AND PEDIA PROBE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
<b>Submission of Bids (Bid Opening): OCTOBER 19, 2022 AT 9:00 AM</b>	<b>Submission of Bids (Bid Opening): OCTOBER 24, 2022 AT 9:00 AM</b>
•Charging Time: 5 hours	Charging Time: 5 hours or better
•100 hours data memory	At least 100 hours data memory
•Trend: 100 hours	Trend: At least 100 hours
REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 – 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 24<sup>th</sup> day of October 2022, 9:00 AM:


This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17<sup>th</sup> day of October 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



# METRO MANILA CENTER FOR HEALTH DEVELOPMENT


Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

## TECHNICAL SPECIFICATIONS

Item No. 1	<b>PULSE OXIMETER WITH ADULT AND PEDIA PROBE</b>	Qty./Unit	<b>17 Units</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>1,190,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>Technical Specifications:</b> <ul style="list-style-type: none"> <li>•Type: Heavy Duty (Tabletop)</li> <li>•Dimensions: Manufacturer's Standard</li> <li>•Power Supply: Dual Power Operation (220V, 60Hz or Battery should be Rechargeable)</li> <li>•Battery Type: Manufacturer's Standard</li> <li>•Operation: At least 4 hours</li> <li>•Charging Time: 5 hours or better</li> <li>•Safety Standard: IEC 60601-1</li> <li>•Electric Shock Protection Standard: MDD 93/42/EEC: Class II b</li> <li>•SpO2 Measuring range: 0 to 100%</li> <li>•SpO2 Accuracy <math>\pm 2</math> (70%- 100%)</li> <li>•At least 100 hours data memory</li> <li>•Includes adult and pediatric reusable sensors/probe</li> <li>•Display: LCD or better</li> <li>•Trend: At least 100 hours</li> <li>•Audible and visible alarm</li> </ul>			



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

**REQUIREMENTS IF AWARDED THE CONTRACT:**

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within **30 – 60** calendar days upon receipt of Notice to Proceed.
2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
3. **Training:** The supplier shall provide training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.
4. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
  - a) Service manual in English language
  - b) Operation manual in English language
7. 'With "**DOH-MMCHD HFEP**"(Government Property not for sale) sticker in each unit

[illegible]

Source of Fund: SAA 2022-02-0531

Recipient/s:

Marikina - 9 units

Muntinlupa - 8 units

**ADDITIONAL REQUIREMENT TO BE SUBMITTED BY THE SINGLE/LOWEST CALCULATED BIDDER (SCB/LCB) AS PART OF POST QUALIFICATION:**



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	
---	--