



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 138E

PROCUREMENT OF 6 UNITS WHEELED DETACHABLE STRETCHER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Submission of Bids (Bid Opening): OCTOBER 19, 2022 AT 9:00 AM	Submission of Bids (Bid Opening): OCTOBER 24, 2022 AT 9:00 AM
Frame: Aluminum Alloy	Frame: Aluminum Alloy - retained
Has a seperable stretcher that can be removed and be used as a separate stretcher	Has a separable stretcher that can be removed and be used as a separate stretcher
Max Load: at least 160kg	Capacity Load: at least 160kg
Dimensions: L x W x H = 2000mm x 600mm x 850mm, ± 50mm in every dimension, adjustable height must be at least in range or bigger	Dimensions: Manufacturer Standard
Adjustable height: 300mm to 850mm	Adjustable height: Manufacturer Standard
REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed	REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 – 60 calendar days upon receipt of Notice to Proceed
Recipients: Muntinlupa – 3 units Municipality of Pateros – 2 units	Recipients: Muntinlupa – 3 units Municipality of Pateros – 3 units

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 24th day of October 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17th day of October 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	WHEELED DETACHABLE STRETCHER	Qty./Unit	6 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 420,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: Frame: Aluminum Alloy Has a separable stretcher that can be removed and be used as a separate stretcher Mattress: At least 3 cm thick, water proof seamless mattress Has 3 belts Castor wheels: At least 100 mm diameter castors Capacity Load: at least 160kg Dimensions: Manufacturer Standard Adjustable height: Manufacturer Standard			
B. REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 – 60 calendar days upon receipt of Notice to Proceed 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

6. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit.

[illegible]

Source of Fund: SAA 2022-02-0531 (HFEP 2022)

Recipients:

Muntinlupa — 3 units

Municipality of Pateros — 3 units

C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.