



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2022 – 063**

**PROCUREMENT OF 1,501 KITS OF COVID ANTIGEN RAPID TEST KIT**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

**Revision and clarification to provisions/specifications in the Bidding Documents:**

\*No changes as stated in technical specifications.


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 11<sup>th</sup> day of May 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 4<sup>th</sup> day of May 2022 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
<b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>PROCUREMENT OF 1,501 KITS OF COVID ANTIGEN RAPID TEST KIT</b>	Qty./Unit	1,501 kits
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>P 600,400.00</b>			
<b>PURCHASER'S SPECIFICATION</b>		<b>STATEMENT OF COMPLIANCE</b>	
<p><b>A. Specifications:</b></p> <p>Technical Specifications:            *Specificity: 90-98%            *Sensitivity: ≥ 90%            *Testing Method: Nasopharyngeal Swabbing (Point of Care)            *Test Time: 15-20 minutes            *Kit content:            *25 Test Devices individually pouched with dessicant, 25 Extraction Tubes, 25 Extraction Cap,            *1 kit with positive and negative controls            *25 Sterilized Nasopharyngeal Swabs for Sample Collection            *1 Sample Tube Rack with self-contained tube with cap for specimen containment using break-off swab            *1 Quick Reference Guide, 1 Insertion for Use            *Valid FDA Certification            *DOH Recommended            *Correlation study with RT-PCR Test Internal of Third Party including but not limited to RITM, NIH, FDA, WHO</p>			
<p><b>B. Upon delivery the following must be complied:</b></p> <p><b>Shelf Life:</b> Supplies must be fresh commercial stock with a total shelf life of at least fifteen (15) months from the date of delivery</p> <p><b>Packaging Instruction:</b> Twenty-five (25) test devices in a kit. Number of kit per carton will follow standard packaging of manufacturer.</p> <p><b>Transport and Storage Requirement</b> The storage and working temperature is between 2-30°C</p> <p><b>Labelling Instructions:</b> *Each box should be imprinted or stickered with non-removable or permanent sticker or label that is binding and with residue and ripping if removed:</p> <p style="text-align: center;"><b>Philippine Government Property-Department of Health-MMCHD Not For Sale</b></p> <p>Date of Manufacture: _____            Date of Expiry: _____            Batch/ Lot No. _____</p> <p>Delivery period: 60 days after receipt of PO            Place of Delivery: MMCHD Pasig Warehouse</p>			
<p><b>C. Additional requirement to be submitted by the Single/ Lowest Calculated Bidder (SCB/ LCB) as part of post-</b></p>			



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**qualification:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.