



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2022 – 121E
PROCUREMENT OF VARIOUS DENTAL EQUIPMENT
(LOT BIDDING)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 1: 2 UNITS DENTAL UNIT, MOTORIZED WITH ACCESSORIES *No changes as stipulated technical specifications.	
Item No. 2: 2 UNITS DENTAL AEROSOL SUCTION MACHINE *No changes as stipulated technical specifications.	
Item No. 3: 16 UNITS DENTAL ULTRASONIC SCALER MACHINE *No changes as stipulated technical specifications.	
FROM	TO
Item No. 4: 21 UNITS STERILIZING UNIT Sterilizing Temperature: 40°C, 125°C, -5°C	Sterilizing Temperature: 40°C, 125°C

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 30th day of August 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 22nd day of August 2022 in MMCHD

Approved by:

SGD.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	DENTAL UNIT, MOTORIZED WITH ACCESSORIES	Qty./Unit	2 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 360,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> · Fully motorized with pre-set functions · Touch control chair · Slim ergonomic design upholstery · LED light with sensor switch · Double side armrest · Colored glass/ ceramic bowl, rotating detachable cuspidor bowl (with plastic strainer) · Two (2) pcs. 3way syringe · Three (3) pcs hand piece attachments (2 hi-speed adaptors & 1 low-speed adaptor) · Auto-cup-filler & bowl flush (water flushing control valve, cup-filler with control valve) · Saliva ejector-built-in motor pump type, does not use air or water · Clean water system · LED X-ray film viewer (Peri-apical X-ray viewer) · Multi-functional foot control · Water tank for distilled water for hand piece · Two (2) hi-speed hand piece · One (1) low-speed complete with straight hand piece & Contra-angle · LED light cure unit (external) · Tissue box with tray and cup dispenser · Silent type and oil-free dental air compressor · Saddle-type with backrest dentist stool 			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 2	DENTAL AEROSOL SUCTION MACHINE	Qty./Unit	2 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 120,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
Technical Specifications: <ul style="list-style-type: none"> ● Super Bio Filter: 0.3 microns HEPA Filter ● With UV light inside the chamber ● Filter change alarm function ● Noise: ≤ 65 dB ● Digital display controls ● Triple Filter System <ul style="list-style-type: none"> - Aerosol Filter - Moisture Contaminant and Dust Filter - Clean Air Filter ● Material: Powder-coated steel ● Airflow: At least 4000 L/min ● Suction Pressure: 25 kpa ● Power Supply: 220V, 60 Hz 			
<u>REQUIREMENTS IF AWARDED THE CONTRACT:</u> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 3	DENTAL ULTRASONIC SCALER MACHINE	Qty./Unit	16 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 448,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Detachable Hand piece with Light (LED), better visibility, autoclavable • Automatic Frequency Tracking, scaling, perio, endo & clinical application • Automatic Water Supply System with Self-Contained Clean water system • Heat free operation • Ten (10) working tips, includes Perio, Scaler and Endodontic • Endodontic Tip attachment • Power Input: 220V- 240V / 50- 60Hz • Output Power: 3W – 20W • Frequency: 28kHz ± 3kHz • Main Unit Weight: manufacturer's Standard • Adapter Weight: Manufacturer's Standard • Dimension: Manufacturer's Standard 			
DOCUMENTARY REQUIREMENTS: <ol style="list-style-type: none"> 1. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language. 2. Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485: Quality Management System – Requirements for regulatory purposes in the name of the manufacturer. The Certificates must be issued by an independent Certifying Body/Agency. 3. Valid Certificate of Distributorship (as first Tier Distributor) issued by the Manufacturer of each equipment authorizing the bidder to sell/distribute the offered equipment. 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>4. Notarized Certificate of the bidder:</p> <p>a. That the brand of the equipment has been in the local and/or international market for at least ten (10) years.</p> <p>b. That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall.</p> <p>5. Bidder's valid and Current License to Operate (LTO) as a medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted: i) Copy of expired LTO, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO.</p>	
<p><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. the supplier shall either repair or replace any item part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft.</p> <p>a) Service manual in English language</p> <p>b) Operation manual in English language</p> <p>6. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit.</p> <p>X-X</p>	



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>Recipient: Muntinlupa City - 16 units</p>	
<p><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	

<p>Republic of the Philippines Department of Health Metro Manila Center for Health Development</p> <p>TECHNICAL SPECIFICATIONS</p>			
Item No. 4	STERILIZING UNIT	Qty./Unit	21 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 367,500.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <p>2-in-1 Dry Heat/Ultraviolet Sterilizer</p> <p>Power Supply: 220V, 50-60Hz</p> <p>Rated Power: 500W</p> <p>Loading Limit: at least 5 kg</p> <p>Sterilizing Temperature: 40°C, 125°C</p> <p>Dimensions: Manufacturer's Standard</p> <p>Stainless Steel Body Frame</p> <p>Three-Layer Tray each Side</p> <p>Build Timer: 30-60 mins (Dry Heat)</p> <p>Can be wall-mounted</p>			



<p>Splash Proof Heater Bulb</p>	
<p>DOCUMENTARY REQUIREMENTS:</p> <ol style="list-style-type: none"> 1. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language. 2. Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485: Quality Management System – Requirements for regulatory purposes in the name of the manufacturer. The Certificates must be issued by an independent Certifying Body/Agency. 3. Valid Certificate of Distributorship (as first Tier Distributor) issued by the Manufacturer of each equipment authorizing the bidder to sell/distribute the offered equipment. 4. Proof (such as sales invoice) that the Brand of the equipment has been sold to other health facilities 5. Notarized Certificate of the bidder: <ol style="list-style-type: none"> a. That the brand of the equipment has been in the local and/or international market for at least ten (10) years. b. That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall. 5. Bidder’s valid and Current License to Operate (LTO) as a medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted: i) Copy of expired LTO, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO. 	
<p><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Warranty: Warranty certificate for two (2) years on parts and service. the supplier shall either repair or replace any part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning 4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period 	

