



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2022 – 153E**

**PROCUREMENT OF 6 UNITS SYRINGE PUMP**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
• Has a rechargeable battery with charger	• Has a rechargeable battery with charger - retained
• With back up power of at least 15-20 hours	• With back up power of at least 15-20 hours - retained
<b>REQUIREMENTS, if awarded the contract</b> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	<b>REQUIREMENTS, if awarded the contract</b> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 5<sup>th</sup> day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 25<sup>th</sup> day of November 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**


Director III / BAC Chairperson

4



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No. 1</b>	<b>SYRINGE PUMP</b>	<b>Qty./Unit</b>	<b>6 Units</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>1,140,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>Technical Specifications:</b> <ul style="list-style-type: none"> <li>Automatic syringe size and model detection</li> <li>Large format LCD/TFT display</li> <li>A minimum flow rate range from 0.1 - 1200 ml/hr for 50ml syringe, 100ml/hr for 20ml syringe and 0.1 - 60 ml/hr for 10ml syringe.</li> <li>Syringe range from 20-50/60 ml</li> <li>A flow rate accuracy of <math>\pm 2\%</math></li> <li>At least 3 levels of programmable occlusion pressure</li> <li>Has a rechargeable battery with charger</li> <li>With back up power of at least 15-20 hours</li> <li>Work with input 100 to 240VAC, 50/60Hz supply</li> </ul>			
<b>REQUIREMENTS, if awarded the contract</b> <ol style="list-style-type: none"> <li><b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.</li> <li><b>2. Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li><b>3. Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance</li> </ol>			



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>staff within 3 days upon delivery of the equipment.</p> <p>4. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. <b>Notarized undertaking</b> that the supplier shall conduct the necessary corrective maintenance, replacements and repair within seven (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"><li>a) Service manual in English language</li><li>b) Operation manual in English language</li></ul> <p>7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</p> <p>X-X</p> <p>Source of Fund: SAA 2022-07-3500</p>	
<p><b><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></b></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	