



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 005

PROCUREMENT OF WATER TESTING KIT WITH COMPARATOR (SHORT OF AWARD)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Packaging - 20 test per kit	
Delivery Area – Pasig Warehouse	
Delivery Period – 30 to 60 calendar days	
Water Testing Kit should be validated by NRL EAMC	
FROM	TO

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 28th day of November 2022, 9:30 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 18th day of November 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	PROCUREMENT OF WATER TESTING KIT WITH COMPARATOR	Qty./Unit	200 KITS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: P 1,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><u>CONTENTS/SPECIFICATIONS:</u></p> <p>A. Coliform Test Reagent</p> <ul style="list-style-type: none"> -Utilizes MUG/ONPG enzyme-substrate test (qualitative) -Can also be used for MPN method (quantitative) -Sensitivity: 1 chlorine-injured coliform per 100 mL sample -Testing time: 16-48 hours Storage: 2-30 °C <p>B. Water Sampling Container</p> <ul style="list-style-type: none"> -Sterile flexible plastic -Water sample container with screw top cover (with sodium thiosulfate) -Capacity: 120mL -With 1 comparator per 200 pieces of sample container <p>Packaging - 20 test per kit</p> <p>Delivery Area – Pasig Warehouse</p> <p>Delivery Period – 30 to 60 calendar days</p> <p>Water Testing Kit should be validated by NRL EAMC</p>			
<p><u>ADDITIONAL REQUIREMENT TO BE SUBMITTED BY THE SINGLE/LOWEST CALCULATED BIDDER (SCB/LCB) AS PART OF POST QUALIFICATION:</u></p> <p>1. One (1) original sample of manufacturer's product to be</p>			



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	
--	--