Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/BID BULLETIN NO. 1

IB 2023 – 013 PROCUREMENT OF CD4 CARTRIDGE, 100 TESTS/BOX (BD FACS PRESTO) (SHORT OF AWARD)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Delivery Site: DOH-MMCHD Pasig Warehouse, Caruncho Avenue, Pasig City	
FROM	ТО
*Individually pouched, POCT (Point Care Test)	*Individually pouched, POCT (Point Care Test)
single use cartridge' Sealed cartridge with no	single use cartridge' Sealed cartridge with no
manual sample handling or processing and with	manual sample handling or processing and with
minimum testing time: 100 cartridges/box;	minimum testing time: 100 cartridges/box;
compatible with existing CD4 analyzer at	compatible with existing CD4 analyzer at
recepient account (BD FACS)	recipient account (BD FACS analyzer)

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 28th day of November 2022, 9:30 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 18th day of November 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH

Director III / BAC Chairperson 😽



Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS PROCUREMENT OF CD4 CARTRIDGE, 100 Item Qty./Unit 13 CARTRIDGE TESTS/BOX (BD FACS PRESTO) (SHORT OF No. 1 AWARD) Name of Manufacturer: · Country of Origin Brand: Model: (if applicable) ABC: P 1,690,000.00 PURCHASER'S SPECIFICATION STATEMENT OF COMPLIANCE **SPECIFICATIONS:** *Standard Features* *Individually pouched, POCT (Point Care Test) single use cartridge' Sealed cartridge with no manual sample handling or processing and with minimum testing time: 100 cartridges/box; compatible with existing CD4 analyzer at recipient account (BD FACS analyzer) Delivery Time: 60 days after receipt of PO Delivery Site: DOH-MMCHD Pasig Warehouse, Caruncho Avenue, Pasig City ADDITIONAL REQUIREMENT TO BE SUBMITTED BY THE SINGLE/LOWEST CALCULATED BIDDER (SCB/LCB) AS PART OF POST QUALIFICATION: 1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.