



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 043E
PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
> One (1) computer set with printer	> One (1) computer set with printer - Processor: at least A6 7480 Motherboard: FM2 Ram: At least 8GB RAM DDR3 HardDisk: At least 256GB SSD Accessories/Peripherals: Keyboard, Mouse, AVR and at least 21-inch 1080p Monitor with HDMI
Remove: Must be compatible with the existing reagents and existing system	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25th day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of May 2023 in MMCHD

Approved by:

SGD
PRETCHELL P. TOLENTINO, MD, MCHM,
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	HEMATOLOGY ANALYZER	Qty./Unit	1 unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 1,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><u>A. SPECIFICATIONS:</u></p> <ul style="list-style-type: none"> • System: Fully Automated 5-part differential • Measurement Principle: Fluorescence staining and Flow cytometry • Measurement Parameters: 28 parameters, including 4 research parameters • Throughput: At least 60 tests/hour • Test Mode: CBC/CBC-Diff • Sample Type: Whole blood/Capillary Blood/Pre-dilution blood • Sampling Method: Manual Sampling • Sample Volume: 20ul • Sample pre-dilution: Manual injection/Automatic injection • Reagents: 4 routine reagents (Diluent, Lysing reagent, Hemolysing reagent, Flourescent Dye) 1 cleaning solution • Quality Management <ul style="list-style-type: none"> - Calibrator should be FDA registered to ensure safety and quality of the machine. • Safety standards: <ul style="list-style-type: none"> - SJ/T 11363-2006 for the toxic and harmful substances of homogeneous materials - IEC 61326 for tested electrical level • Data Storage: At least 200,000 data • Inclusion: <ul style="list-style-type: none"> - One (1) set of start-up reagents and accessories with GHL (Glucose, Hemoglobin and Lactate) meter kit - One (1) computer set with printer - <p>Processor: at least A6 7480 Motherboard: FM2 Ram: At least 8GB RAM DDR3 HardDisk: At least 256GB SSD Accessories/Peripherals: Keyboard, Mouse, AVR and at least 21-inch 1080p Monitor with HDMI - One (1) unit of at least 1KVA UPS</p>			
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none">a) Service manual in English languageb) Operation manual in English language <p>7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit</p> <p>x-x</p> <p>Source of Fund: SAA 2022-08-3617 (ConAp 2022)</p> <p>Recipients:</p> <p>Health Facilities in Marikina City – 1 unit</p>	
--	--