



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 046E

PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Add: >Protective Lens for light cure machine >1 for Pedia >1 for Dentist	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25th day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of May 2023 in MMCHD

Approved by:

SGD

PRETCHELL P. TOLENTINO, MD, MCHM,

Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	PEDIATRIC DENTAL CHAIR UNIT WITH ACCESSORIES	Qty./Unit	1 unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 300,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><u>A. SPECIFICATIONS:</u></p> <ul style="list-style-type: none"> • Special design for children • All plastics are made of injection mold • Normal operating light • Normal suction system • Three-way syringe (One for hot and one for cold), 2 pcs • Built-in water purified system • Automatic water heating system • Built-in X-ray film viewer • With 2 pieces of dental stool (1 saddle type) • Inclusions <ul style="list-style-type: none"> - 1 unit of Intra Oral Camera with at least 17" LED monitor - 1 unit of Ultrasonic Scaler with LED (with 4 tips included) - 1 unit of Light Cure Machine - 3 pieces of Handpiece 4 holes or 2 holes - 2 hp Compressor with at least 25L capacity and 115 psi of max. Pressure - Protective Lens for light cure machine -1 for Pedia -1 for Dentist 			
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The</p>			

