



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2023 – 050E**

### PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
Delivery: >30 calendar days	Delivery: > 30-60 calendar days.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19<sup>th</sup> day of May 2023 in MMCHD

Approved by:

SGD

**PRETCHELL P. TOLENTINO, MD, MCHM,**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<p><b>Republic of the Philippines</b></p> <p><b>Department of Health</b></p> <p><b>Metro Manila Center for Health Development</b></p> <p><b>TECHNICAL SPECIFICATIONS</b></p>			
Item No. 1	<b>LABORATORY OVEN (90L)</b>	Qty./Unit	4 units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
<b>ABC: 800,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><b><u>A. SPECIFICATIONS:</u></b></p> <ul style="list-style-type: none"> <li>• Application:               <ul style="list-style-type: none"> <li>- Annealing</li> <li>- Die-bond curing</li> <li>- Drying</li> <li>- Polyimide baking</li> <li>- Sterilizing</li> </ul> </li> <li>• Temperature:               <ul style="list-style-type: none"> <li>- Range: Ambient Temp. +5 °C up to 250 °C</li> <li>- Accuracy: ±1.0 °C</li> <li>- Has even temperature distribution</li> </ul> </li> <li>• Construction               <ul style="list-style-type: none"> <li>- Capacity: At least 90 L</li> <li>- Number Trays: At least 2</li> <li>- Tray Material: Stainless wire mesh</li> <li>- Controller: Microcontroller for internal temperature</li> <li>- Heater: U shaped Stainless Steel with Nichrome wire air heater</li> <li>- Insulation: High Density Ceramic Glass Wool or better</li> <li>- Interior: Stainless Steel 304</li> <li>- Exterior: Powder Coated Steel</li> </ul> </li> <li>• Power Supply: 220V, 50/60 Hz</li> </ul>			
<p><b><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></b></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <i>30-60 calendar days</i> upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p>			

