



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

#### IB 2022 – 060E PROCUREMENT OF 1 UNIT COLONOSCOPE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Usage: Adult	
• Must be compatible with the existing equipment of the hospital	
• Video Processor: PentaX EPK-1000	
FROM	TO
• Working Length: 1,700	• Working Length: 1,700 mm
• Angulation: 180/180	• Angulation: 180/180 up and down
• Angulation: 160/160	• Angulation: 160/160 right and left

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23<sup>rd</sup> day of May 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of May 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

## Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>COLONOSCOPE</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,560,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<b>A. Technical Specifications:</b> <ul style="list-style-type: none"> <li>• Channel Diameter: maximum of 3.8 mm</li> <li>• Insertion Tube Diameter: maximum of 11.6 mm</li> <li>• Working Length: 1,700 mm</li> <li>• Angulation: 180/180 up and down</li> <li>• Angulation: 160/160 right and left</li> <li>• Angle of view: 140°</li> <li>• Product life: Minimum of 4-5 years</li> </ul> Usage: Adult <ul style="list-style-type: none"> <li>• Must be compatible with the existing equipment of the hospital</li> <li>• Video Processor: PentaX EPK-1000</li> </ul>			
<b>B. REQUIREMENTS IF AWARDED THE CONTRACT:</b> <ol style="list-style-type: none"> <li>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</li> <li>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment</li> <li>4. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and</li> </ol>			

