

Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 072E

### PROCUREMENT OF 1 UNIT HYPERTHERMIA, BLOOD FLUID WARMING

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>
--

*No changes as stated in the technical specifications.
--

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23<sup>rd</sup> day of May 2022, 9:00 AM:


This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of May 2022 in MMCHD

Approved by:


**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

## Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
<b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>HYPERTHERMIA, BLOOD FLUID WARMING</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>270,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b><u>TECHNICAL SPECIFICATIONS:</u></b> <ul style="list-style-type: none"> <li>With two airflow settings and hose-end temperature sensor</li> <li>Capable of recording machine errors and therapy reports</li> <li>With built-in hour meter or service meter</li> <li>Dimensions: Manufacturer's Standard</li> <li>Noise Level: <math>\leq 55</math> dBA</li> <li>Hose: Flexible</li> <li>Filtration System: At least 0.2 <math>\mu</math>m air filter</li> <li>Mounting: Can be clamped on to an IV pole, placed on a hard surface or mounted on a rolling stand accessory.</li> <li>Temperature Control: Electronically controlled.</li> <li>Alarm System: With audible and visual alarm</li> <li>Blower Motor               <ul style="list-style-type: none"> <li>- Speed: Max of at least 4500 rpm for high-fan setting</li> <li>- Airflow: At least up to 45 cfm</li> </ul> </li> <li>Power Supply: 220V, 60 Hz</li> <li>Inclusion:               <ul style="list-style-type: none"> <li>- Rolling Cart</li> </ul> </li> </ul>			
<b><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></b> <ol style="list-style-type: none"> <li><b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</li> <li><b>2. Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li><b>3. Training:</b> The supplier shall provide a training on the proper use and</li> </ol>			



1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.