

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 073

PROCUREMENT OF 250 UNITS STADIOMETER AND INFANTOMETER MAT

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Material – ABS Plastic	
FROM	TO
ABC: P 3,600,000.00	ABC: P 3,625,000.00
Stadiometer: Measuring range 20-205cm/8-18”	Stadiometer: Measuring range 20-205cm/8-81”
Infantometer Mat Dimensions: (WxHxD): 300 mm x 140 mm x 1340 mm	Infantometer Mat Dimensions: (WxHxD): 300 mm x 140 mm x 1250mm to 1350 mm
Weight: approx. 500g	Weight: at least 500g to 750g


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 19th day of July 2022, 9:00 AM:


This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 12th day of July 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson




METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	PROCUREMENT OF 250 UNITS STADIOMETER AND INFANTOMETER MAT	Qty./Unit	250 UNITS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: P 3,625,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><u>SPECIFICATIONS:</u></p> <p>Stadiometer:</p> <p>Simple and easy to set up</p> <p>No wall fastening necessary</p> <p>Stable large floor plate ensures stability</p> <p>Keeps the rod straight and stable</p> <p>Result clearly visible while measuring</p> <p>Convenient and easy to transport</p> <p>Measuring range 20-205cm/8-81"</p> <p>Graduation: 1mm, 1mm/1/8</p> <p>Dimension: (WxHxD): 337 x 2.130 x 590mm / 13.3 x 177 x 624/mm 13.3 x 7 x 24.6"</p> <p>Device Weight: 2.4 kg/5.3 lbs</p> <p>Material – ABS Plastic</p> <p>Infantometer Mat</p> <p>Light-weight and easy to clean</p> <p>Suitable for commercially available disinfectants</p> <p>Fixed head piece and the sliding foot positioner make it simple to use</p> <p>Measuring mat rolls up for easy transport and space-saving storage</p> <p>Measurement range: 100-990mm/4-39 inches</p> <p>Graduations: 5mm/ 1/4 inch</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>Accuracy: greater than 5mm</p> <p>Dimensions: (WxHxD): 300 mm x 140 mm x 1250mm to 1350 mm</p> <p>Weight: at least 500g to 750g</p> <p>Temperature range: +10 °C to 40 °C</p>	
<p><u>UPON DELIVERY THE FOLLOWING MUST BE COMPLIED:</u></p> <p>Labelling Instructions</p> <p>On each tape, box, carton should be legibly imprinted or stickered with non-removable or permanent sticker/label that is binding, and with residue and tearing if removed:</p> <p>Philippine Government Property- Department of Health</p> <p>NOT FOR SALE</p>	
<p><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of pos-qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	