

## Republic of the Philippines Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

## SUPPLEMENTAL/BID BULLETIN NO. 1

## IB 2022 – 073E PROCUREMENT OF 5 UNITS FETAL DOPPLER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:				
FROM	TO			
• Power Supply: 220VAC, 60 Hz or Battery operated	• Power Supply: 220VAC, 60 Hz			

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23<sup>rd</sup> day of May 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15th day of May 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH

Director III / BAC Chairperson



# Republic of the Philippines Department of Health METRO MANILA CENTER FOR HEALTH DEVELOPMENT

# **Technical Specifications**

Republic of the Philippines				
Department of Health				
Metro Manila Center for Health Development .				
TECHNICAL SPECIFICATIONS				
Item No. 1	FETAL DOPPLER	Qty./Unit	5 Units	
Name of Manufacturer:			Country of Origin	
			Model: (if applicable)	
Brand:			Model. (II applicable)	
ABC: <b>225,000.00</b> PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
TECHNICAL SPECIFICATIONS:				
Power Supply: 220VAC, 60 Hz				
Type: Table Top (heavy duty)				
Material: Made of hard plastic/ ABS Plastic				
Standard Features:			5	
- LCD Display				
- Fetal Hea	rt Rate			
- Built-in loudspeaker: with noise reduction				
- Probe holder				
<ul><li>Standa</li></ul>	rd Accessories:			
- 1 conductive gel				
- 1 ultrasor	nic probe			
REQUIREMENTS IF AWARDED THE CONTRACT:				
1. <b>Completion Period</b> : The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.				
2. <b>Testing</b> : Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.				
<b>3. Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment				
<b>4.Warranty</b> : Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the				



## Republic of the Philippines Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

- 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period
- 6. **Manuals**: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
- a) Service manual in English language

Recipient: Ospital ng Tondo

C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.