



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	TRANSPORT INCUBATOR	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 380,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Air mode and skin mode controlled by micro-computer • Power <ul style="list-style-type: none"> - Use AC Power and DC Power alternatively, DC Power on the ambulance - On mains 220, 60Hz • Battery: <ul style="list-style-type: none"> - has LED indicator for battery power capacity, battery charge and battery - has heater power condition (25 to 100%) - is maintenance free - is rechargeable - should support all function together for at least 2 hours • Set temperature, air temperature, skin temperature • Self-testing function, various failure alarms by audible and visual • Alarm system: <ul style="list-style-type: none"> - Alarm indicators for power failure, high temperature deviation, fan failure, temperature sensor failure, power failure, low DC • Double wall hood with side door and front door, 4 operating windows and 2 iris ports, 5 infusion seals • The infant bed can be pulled out from side door easily • Natural air flow humidity 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

- has oxygen supply system
- has LED observation/examination light, adjustable brightness
- has built-in storage cabinet
- Sturdy Trolley
 - Collapsible trolley with provision to keep re-fillable oxygen cylinder and battery
 - the trolley must be sturdy with sturdy and shock absorbing wheels
 - has space for accessories
- Should have <60dBA noise level
- Should have O2 concentration ranging 21% to 58% minimum
- Should have additional features for accessory shelf, IV pole, high hood, pressure regulator and flow meter.

REQUIREMENTS IF AWARDED THE CONTRACT:

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.


2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.

3. **Training:** The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within (3) days upon delivery of the equipment.

4. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

5. **Notarized undertaking** that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2022 – 101E
PROCUREMENT OF 1 UNIT TRANSPORT INCUBATOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
• Should have optional features for accessory shelf, IV pole, high hood, pressure regulator and flow meter.	Should have additional features for accessory shelf or bag, IV pole, high hood, pressure regulator with flow meter

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 27th day of June 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of June 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson