



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 163E

PROCUREMENT OF 30 UNITS FETAL DOPPLER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Add: ● Standard Accessories: -1 carry case	
FROM	TO
- LCD Display	LCD Display or better LCD Display: at least 80mm
REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.	REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 13th day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6th day of December 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	FETAL DOPPLER	Qty./Unit	30 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 1,350,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> ● Power Supply: Dual Power, (220VAC/60 Hz and Battery operated) ● Type: Table Top (heavy duty) ● Material: Made of hard plastic/ ABS Plastic ● Standard Features: <ul style="list-style-type: none"> - LCD Display or better - LCD Display: at least 80mm - Fetal Heart Rate - Built-in loudspeaker: with noise reduction - Probe Holder ● Standard Accessories: <ul style="list-style-type: none"> - 1 gallon or 4 liter conductive cream/gel - 1 ultrasonic probe - 1 carry case 			
REQUIREMENTS IF AWARDED THE CONTRACT: <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 to 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning 			

