



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 164E

PROCUREMENT OF 25 UNITS VITAL SIGNS MONITOR (SPO2, BP AND TEMPERATURE)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
• NIBP (Non-invasive Blood Pressure) REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	• NIBP (Non-invasive Blood Pressure) REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30-60 calendar days upon receipt of Notice to Proceed.


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 13th day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6th day of December 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	VITAL SIGNS MONITOR (SPO2, BP AND TEMPERATURE)	Qty./Unit	25 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,750,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: Type: Handheld <ul style="list-style-type: none"> • Blood pressure rate, Oxygen saturation, Pulse rate, Pulse intensity, Temperature • Easy operation by 3.5" color high lightness LCD display lightness is adjustable • Resolution: 320*480 • Anti-motion artifacts, anti-perfusion • Unique Large Font and Waveform display • Dynamic data and Graphic trends to view • Audio and Visual Alarm • Applies to adults, children and newborn • Trend Storage: 72 hours • Alarm: Probe off, Finger out, Low power • NIBP (Non-invasive Blood Pressure) -Measuring range:			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Adult:	PED:	NEO:
SYS: 40 ~ 270mmHg	SYS: 40 ~ 200mmHg	SYS: 40 ~ 135mmHg
MEAN: 20 ~ 230mmHg	MEAN: 20 ~ 165mmHg	MEAN: 20 ~ 105mmHg
DIA: 10 ~ 210mmHg	DIA: 10 ~ 150mmHg	DIA: 10 ~ 95mmHg
PULSE RATE: 40~240bpm	PULSE RATE: 40~240bpm	PULSE RATE: 40~240bpm
<ul style="list-style-type: none"> -Blood pressure precision average deviation: $\leq \pm 5$mmHg -Blood pressure precision standard deviation: ≤ 8mmHg 		
<ul style="list-style-type: none"> • SPO2 Pulse Oxygen Saturation: <ul style="list-style-type: none"> - Measurement: Absorption of a spectral -Resolution: 1% -Accuracy: ± 2bpm (non-motion) and ± 3bpm (motion) -Pulse rate: 30-250 bpm • Temperature: <ul style="list-style-type: none"> - Range: 25 - 45°C - Resolution: 0.1°C - Sensitivity: ± 0.1°C • Power: <ul style="list-style-type: none"> - Lithium Battery, operation time at least 15 hours for normal operation - 100-240VAC, 50/60Hz, 30VA - Input fuse: 2A/250V 		
<p>REQUIREMENTS IF AWARDED THE CONTRACT:</p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30-60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment. 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship 		

