



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 171E

PROCUREMENT OF 10 UNITS CLINICAL CENTRIFUGE, 12 PLACER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:
--

Add:

Type: Digital

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 13th day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6th day of December 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "A. Sudiacal", is written over the printed name.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	CLINICAL CENTRIFUGE, 12 PLACER	Qty./Unit	10 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 800,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
TECHNICAL SPECIFICATIONS: Type: Digital Material: Strong Plastic Case with Metal Lid Brushless motor LED Display speed and timer Alarm: overspeed and imbalance With mechanical door lock and safety switch Error code display speed up and speed down adjustment suitable for qualitative analysis of blood serum and plasma Max RPM: 4000 RPM or Higher Max RCF: at least 2300 x g Timer: 0 to 90 minutes Rotor Capacity: 12 by 20 ml/15 ml/10 ml/5 ml Power: 220V 50/60 Hz Dimension: Manufacturer's Standard			
REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30			

