



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB NO. 2023-053E**  
**PROCUREMENT OF 4 UNITS DENTAL UNIT, MOTORIZED WITH**  
**ACCESSORIES**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
No changes as stipulated in technical specification	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 15, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9<sup>th</sup> day of June, 2023 in MMCHD.

Approved by:

SGD.  
**PRETCHELL P. TOLENTINO, MD, MCHM**  
Director III / BAC Chairperson

**Section VII. Technical Specifications**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>DENTAL UNIT, MOTORIZED WITH ACCESSORIES</b>	Qty./Unit	<b>4 UNITS</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>720,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> <li>· Fully motorized with pre-set functions</li> <li>· Touch control chair</li> <li>· Slim ergonomic design upholstery</li> <li>· LED light with sensor switch</li> <li>· Double side armrest</li> <li>· Colored glass/ ceramic bowl, rotating detachable cuspidor bowl (with plastic strainer)</li> <li>· Two (2) pcs. 3 way syringe</li> <li>· Three (3) pcs hand piece attachments (2 hi-speed adaptors &amp; 1 low-speed adaptor)</li> <li>· Auto-cup-filler &amp; bowl flush (water flushing control valve, cup-filler with control valve)</li> <li>· Saliva ejector-built-in motor pump type, does not use air or water</li> <li>· Clean water system</li> <li>· LED X-ray film viewer (Peri-apical X-ray viewer)</li> <li>· Multi-functional foot control</li> <li>· Water tank for distilled water for hand piece</li> <li>· Two (2) hi-speed hand piece</li> <li>· One (1) low-speed complete with straight hand piece &amp; Contra-angle</li> </ul>			



**Section VI. Schedule of Requirements**

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

<b>Description</b>	<b>Quantity</b>	<b>Delivery Site</b>	<b>Delivery Period</b>
DENTAL UNIT, MOTORIZED WITH ACCESSORIES	4UNITS	Health Facility in National Capital Region	30 calendar days upon receipt of Notice to proceed.

Name of Company:\_\_\_\_\_

Address:\_\_\_\_\_

Signature Over Printed Name :\_\_\_\_\_

Telephone/Fax Number :\_\_\_\_\_

Email:\_\_\_\_\_