



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB NO. 2023-058E**  
**PROCUREMENT OF 7 UNITS VACCINE REFRIGERATOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
> Stainless steel Handle	>- Stainless steel Handle or ABS Plastic – <b>RETAIN</b>
> Refrigerant: R-134a (CFC-free)	> Refrigerant: R-134a <del>or its equivalent</del> (CFC-free) - <b>RETAIN</b>
REQUIREMENTS, if awarded the contract 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	REQUIREMENTS, if awarded the contract 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <u>30-60</u> calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 15, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9<sup>th</sup> day of June, 2023 in MMCHD.

Approved by:

SGD.  
**PRETCHELL P. TOLENTINO, MD, MCHM**  
Director III / BAC Chairperson

**Section VII. Technical Specifications**

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

**TECHNICAL SPECIFICATIONS**

Item No. <b>1</b>	<b>VACCINE REFRIGERATOR</b>	Qty./Unit	<b>7 UNIT</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>1,680,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><b>TECHNICAL SPECIFICATION</b></p> <ul style="list-style-type: none"> <li>• Capacity: at least 80L</li> <li>• Construction: <ul style="list-style-type: none"> <li>- Stainless steel Handle</li> <li>- Drop-proof Polyethylene cabinet or better</li> <li>- Silicone seal</li> </ul> </li> <li>• Alarm: <ul style="list-style-type: none"> <li>- Sensor Error Alarm, High/Low temperature alarm, Controller failure alarm</li> <li>- Power failure alarm</li> </ul> </li> <li>• Power Supply: <ul style="list-style-type: none"> <li>- AC: 220V 50/60 Hz</li> <li>- DC: 12V</li> <li>- Power failure backup 8 hours</li> </ul> </li> <li>• Refrigeration System <ul style="list-style-type: none"> <li>- Refrigerant: R-134a (CFC-free)</li> <li>- Temperature controller: Microprocessor</li> </ul> </li> <li>• Display: Digital</li> <li>• Temperature Accuracy: 0.1° C</li> </ul>			

<ul style="list-style-type: none"><li>• Refrigeration Temperature 2° C ~ 8° C</li><li>• Freezer Temperature: -10° C ~ -25° C</li><li>• Has Data logger included</li><li>• Noise: 46 dB or Better</li><li>• Dimensions: Manufacturer’s Standard</li><li>• Accessories:<ul style="list-style-type: none"><li>- 220V power cord, 12V power cord, Plastic Frames</li></ul></li></ul>	
--	--

<p>REQUIREMENTS, if awarded the contract</p> <ol style="list-style-type: none"><li>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <i>30-60</i> calendar days upon receipt of Notice to Proceed.</li><li>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li><li>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</li><li>4. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li><li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</li><li>6. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:<ol style="list-style-type: none"><li>a) Service manual in English language</li><li>b) Operation manual in English language</li></ol></li><li>7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</li></ol> <p>x-x</p> <p>Source of Fund: SAA 2023-02-000687</p> <p>Recipient: Health Facilities in Marikina City – 7 units</p>	
--	--

*Section VI. Schedule of Requirements*

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

<b>Description</b>	<b>Quantity</b>	<b>Delivery Site</b>	<b>Delivery Period</b>
<b>VACCINE REFRIGERATOR</b>	7units	Health facilities in Marikina City	60 calendar days upon receipt of the Notice to Proceed

Name of Company:\_\_\_\_\_

Address:\_\_\_\_\_

Signature Over Printed Name :\_\_\_\_\_

Telephone/Fax Number :\_\_\_\_\_

Email:\_\_\_\_\_