



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-062E
PROCUREMENT OF 1 UNIT AIR SPRAY PULVERIZATOR MACHINE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
No changes as stipulated in technical specification	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 15, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9th day of June, 2023 in MMCHD.

Approved by:

SGD.
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	AIR SPRAY PULVERIZATOR MACHINE	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC 480,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> • Spray by pulverization composition of Hydrogen Peroxide and silver in the form of particles less than 5 micron • No irritation, no odor, no residue, high material compatibility, environmental friendly, safe for use • Air test strip for disinfection validation • Provides high level disinfection, kills spores at surfaces. • Capable to disinfect room volume from 1 to 999, cubic meters • Disinfection time not more than 3 minutes for 50 cubic meters • The treated room can be re-used in 30-60 minutes • Programmable time to start and time to stop • Automatic printing of disinfection data through printer • Stored in memory up to 100 room data cycle and printable if desired • With RF remote control to start and paused a disinfection cycle • The operation can be paused and resumed • Adjustable screen brightness • Password protected • Visible and audible warning system for start and end of operation. • Portable and transferable with plastic grip • Particle filter keeps the air access system clean 			

- ABS body with 304 grade stainless steel metal parts
- Conforms to LVD and EMC directives, 93/42 EEC Medical Device Directive Class I
- 220-240 VAC, 50/60Hz, 1100 W with 2 meter long power supply cable
- Disinfectant Solution Specification
- Disinfectant solution composed of 6% Hydrogen Peroxide with .003% colloidal silver and other auxiliary substances
- Aldehyde and phenol Free
- 50 pcs monitoring strips
- 5 rolls thermal printer paper
- 5 liters Disinfection Solution
- Packaging in 1 liter opaque bottle

Requirements if awarded the Contract

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.
2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
3. **Training:** The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.
4. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
 - a) Service manual in English language
 - b) Operation manual in English language
7. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
AIR SPRAY PULVERIZATOR MACHINE	1 UNIT	OSPITAL NG MALABON	30 Calendar days upon receipt of the Notice to Proceed.

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____