



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-067E
PROCUREMENT OF 21 UNITS INFUSION PUMP

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
>Horizontal Door Clip	>Horizontal or Vertical Door Clip
>Display: At least 4 inches, LCD with big font	>Display: At least 3 inches, LCD or better
>Micro Infusion Speed: 0.1ml/h up to 1500 ml/h	>Micro Infusion Speed or its equivalent: 0.1ml/h up to <u>999</u> ml/h or better
>Infusion Accuracy: < $\pm 3\%$	>Infusion Accuracy: < $\pm 5\%$
> Power Supply: Built-in battery (manufacturer's standard), can last up to 8 hours of operation if fully charged	> Power Supply: Built-in battery, at least 4 hours of operation or better
Add: With drip sensor	
Requirements if awarded the Contract 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.	Requirements if awarded the Contract 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <u>30-60</u> calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 15, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9th day of June, 2023 in MMCHD.

Approved by:

SGD.
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	INFUSION PUMP	Qty./Unit	21 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC 1,785,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
Technical Specifications: <ul style="list-style-type: none"> • Horizontal <u>or Vertical</u> Door Clip • Display: At least <u>3</u> inches, LCD or better • Ergonomics operation interface • Modes <ul style="list-style-type: none"> - Rate Mode - Drop Mode - Time Mode • Micro Infusion Speed: 0.1ml/h up to <u>999</u> ml/h or better • Infusion Accuracy: < <u>±5%</u> • With drip sensor • With software adjusting function suitable for any brand of infusion sets • With high precision pressure sensor • With built-in drip sensor • Capable of storing up to 800 infusion records and configuration memory • Power Supply: Built-in battery, at least 4 hours of operation or better 			

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
INFUSION PUMP	21UNITS	Health Facilities in national Capital Region	30-60 calendar days upon receipt of Notice to Proceed.

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____