



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB NO. 2023-069E  
PROCUREMENT OF 1 UNIT OPERATING ROOM TABLE (HEAVY DUTY)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
>Alarm type: Sound Alarm	> Alarm type: Sound Alarm – <b>RETAIN</b>
>Accessories <ul style="list-style-type: none"> <li>- Anaesthetist's frame</li> <li>- I.V. stand</li> <li>- Shouder support</li> <li>- Body support</li> <li>- Surgical table arm boards</li> <li>- Arm support</li> <li>- Arm/hand surgical table</li> <li>- Wrist grip</li> <li>- Knee crutches</li> <li>- Arthroscopic knee crutch</li> <li>- Leg rest - Lithotomy position</li> <li>- Foot supports</li> <li>- Foot pedal</li> <li>- X-ray cassette holder</li> <li>- Waste tray</li> <li>- Body strap</li> <li>- Compact stand for surgical table accessories</li>   <li>- Orthopedic traction system</li> <li>- Horseshoe headrest</li> <li>- Skull clamp</li> <li>- Crossbar adaptor for skull clamp</li> <li>- Stand for skull clamp and horseshoe headrest</li> <li>- Carbon fiber leg section</li> <li>- Arthroscopic shoulder support</li> </ul>	>Accessories <ul style="list-style-type: none"> <li>- Anesthetist's frame</li> <li>- I.V. stand</li> <li>- Shoulder support</li> <li>- Body support</li> <li><del>- Surgical table arm boards</del></li> <li>- Arm support with restraining belt/strap</li>   <li>- Arm/hand surgical table</li> <li><del>- Wrist grip</del></li> <li>- Knee crutches with restraining belt/strap</li> <li><del>- Arthroscopic knee crutch</del></li> <li><del>- Leg rest - Lithotomy position</del></li> <li>- Foot supports with clamp</li> <li><del>- Foot pedal</del></li> <li>- X-ray cassette holder</li> <li>- Waste tray</li> <li>- Body strap</li> <li>- Compact stand for surgical table accessories</li> <li>- Orthopedic traction system:</li> <li>- All connecting parts</li> <li>- Traction arms (longitudinally adjustable)</li> <li>- Pair of traction boots (adult)</li> <li>- Traction support rise in conjunction with table</li> <li>- Pair of traction mechanism with fine adjustment</li> <li><del>- Horseshoe headrest</del></li> <li><del>- Skull clamp</del></li> <li><del>- Crossbar adaptor for skull clamp</del></li> <li><del>- Stand for skull clamp and horseshoe headrest</del></li> <li><del>- Carbon fiber leg section</del></li> <li><del>- Arthroscopic shoulder support</del></li> </ul>
Requirements if awarded the Contract  <b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <b>60</b> calendar days upon receipt of Notice to Proceed.	Requirements if awarded the Contract  <b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed

	with <b>60-90</b> calendar days upon receipt of Notice to Proceed.
--	--

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 15, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9<sup>th</sup> day of June, 2023 in MMCHD.

Approved by:

SGD.

**PRETCHELL P. TOLENTINO, MD, MCHM**

Director III / BAC Chairperson

**Section VII. Technical Specifications**

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

**TECHNICAL SPECIFICATIONS**

Item No. <b>1</b>	<b>OPERATING ROOM TABLE (HEAVY DUTY)</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC <b>3,500,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> <li>• Height adjustment: at least 650 - 1000mm</li> <li>• With Mattress: at least 60mm</li> <li>• Power supply: 220V, 50/60Hz</li> <li>• Battery: <ul style="list-style-type: none"> <li>- Maintenance free lead acid or better</li> <li>- Charging time: 18h or better</li> </ul> </li> <li>• Alarm type: Sound Alarm</li> <li>• Alarm: Low battery alarm</li> <li>• Positions: <ul style="list-style-type: none"> <li>- Head section: Upward 45°, Downward 60°, Removable</li> <li>- Backrest section: Upward 80°, Downward 40°</li> <li>- Leg section: Upward 30°, Downward 90°, Spreading 90°, Independent movement, Removable</li> <li>- Lateral: Left 20°, Right 20°</li> <li>- Trendelenburg: 35°</li> </ul> </li> <li>• Tabletop Levelling: parallel to the ground</li> <li>• Accessories <ul style="list-style-type: none"> <li>- Anaesthetist's frame</li> </ul> </li> </ul>			

- I.V. stand
- Shoulder support
- Body support
- Arm support with restraining belt/strap
- Knee crutches with restraining belt/strap
- Foot supports with clamp
- X-ray cassette holder
- Waste tray
- Body strap
- Compact stand for surgical table accessories
- Orthopedic traction system:
  - All connecting parts
  - Traction arms (longitudinally adjustable)
  - Pair of traction boots (adult)
  - Traction support rise in conjunction with table
  - Pair of traction mechanism with fine adjustment

Requirements if awarded the Contract

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with **60-90** calendar days upon receipt of Notice to Proceed.

2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.

3. **Training:** The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.

4. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of



**Section VI. Schedule of Requirements**

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

<b>Description</b>	<b>Quantity</b>	<b>Delivery Site</b>	<b>Delivery Period</b>
<b>OPERATING ROOM TABLE(HEAVY DUTY)</b>	1 unit	Caloocan City Medical Center	60-90 calendar days upon receipt of the Notice to Proceed.

Name of Company: \_\_\_\_\_

Address: \_\_\_\_\_

Signature Over Printed Name : \_\_\_\_\_

Telephone/Fax Number : \_\_\_\_\_

Email: \_\_\_\_\_