



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-073E
PROCUREMENT OF 10 UNITS TRANSFER STRETCHER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
> Dimensions: L x W x H = 2000mm x 700mm x 700mm, \pm 50mm on each	> Dimensions: L x W x H = 2000mm x 700mm x 700mm, \pm 100mm on each dimension

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 15, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9th day of June, 2023 in MMCHD.

Approved by:

SGD.
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	TRANSFER STRETCHER	Qty./Unit	10 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC 700,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Frame: Powder Coated Steel or better • Backraise inclination: at least 0 to 35 degrees • Has adjustable height • Side guard: Made of ABS Plastic or better • Castor wheels: Manufacturer's Standard • Max Load: at least 150kg • Weight: Manufacturer's Standard • Dimensions: L x W x H = 2000mm x 700mm x 700mm, ± <u>100</u>mm on each dimension • Accessories: <ul style="list-style-type: none"> -Has oxygen cylinder holder -Has IV pole holder located in 4 different positions -With one IV pole 			
Requirements if awarded the Contract <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and 			

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
TRANSFER STRETCHER	10 UNITS	Ospital ng Malabon	30 calendar days upon receipt of the Notice to Proceed

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____