



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB NO. 2023-074E**  
**PROCUREMENT OF 9 UNITS FETAL MONITOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
>Fetal Heart Rate Waveform Display Range: 50-210 bpm	> Fetal Heart Rate Waveform Display Range: 50-210 bpm - <b>RETAIN</b>
>Has built-in battery: 4-8 hours	> Has built-in battery: has at least 8 hours or better
>Power aid module: 10-12 hours	> Power aid module: at least 10-12 hours or better
Requirements if awarded the Contract  <b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.	  <b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <b>30-60</b> calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 15, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9<sup>th</sup> day of June, 2023 in MMCHD.

Approved by:

SGD.

**PRETCHELL P. TOLENTINO, MD, MCHM**  
Director III / BAC Chairperson

**Section VII. Technical Specifications**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No. 1</b>	<b>FETAL MONITOR</b>	<b>Qty./Unit</b>	<b>9 Units</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
<b>ABC 1,440,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> <li>• Dual pulsed Doppler FHR</li> <li>• Ultrasound Frequency: 1 MHz</li> <li>• Fetal Heart Rate Measurement Range: 30-240 bpm</li> <li>• Fetal Heart Rate Waveform Display Range: 50-210 bpm</li> <li>• With auto-detection of dual fetal movement</li> <li>• With array-type thermal printer</li> <li>• With automatic printing function</li> <li>• At least 7-inch. TFT color LCD Display</li> <li>• 3-channel display: FHR I, FHR II, UC</li> <li>• With Clinical event marking</li> <li>• At least 160 patient data saving</li> </ul>			



Recipient: Malabon City – 1 unit Municipality of Pateros – 1 unit Caloocan City Medical Center – 3 units Ospital ng Muntinlupa – 2 units Navotas City – 2 units	
--	--

Name of Company: \_\_\_\_\_  
Address: \_\_\_\_\_  
Signature Over Printed Name : \_\_\_\_\_  
Telephone/Fax Number : \_\_\_\_\_  
Email: \_\_\_\_\_

***Section VI. Schedule of Requirements***

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

<b>Description</b>	<b>Quantity</b>	<b>Delivery Site</b>	<b>Delivery Period</b>
<b>FETAL MONITOR</b>	9 UNITS	Health facilities in National Capital Region	30-60 calendar days upon receipt of the Notice to Proceed

Name of Company: \_\_\_\_\_  
Address: \_\_\_\_\_  
Signature Over Printed Name : \_\_\_\_\_  
Telephone/Fax Number : \_\_\_\_\_  
Email: \_\_\_\_\_