



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB NO. 2023-077E
PROCUREMENT OF 2 UNITS INCUBATOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
No changes as stipulated in technical specification	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 15, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9th day of June, 2023 in MMCHD.

Approved by:

SGD.
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	INCUBATOR	Qty./Unit	2 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC 720,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Mode: Air mode and baby mode • Controller: micro-computer • Display: Separate display for set temperature, air temperature and baby temperature • Natural air flow humidity • Adjustable bassinet inclination • Alarm: failure alarm indication 			
Requirements if awarded the Contract <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of 			

acceptance by the end-user after testing and commissioning.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit

x-x

Source of Fund: SAA 2023-02-000687 (HFEP 2023)

Recipient:

Ospital ng Muntinlupa – 2 units

Name of Company: _____
 Address: _____
 Signature Over Printed Name : _____
 Telephone/Fax Number : _____
 Email: _____

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
INCUBATOR	2 UNITS	Ospital ng Muntinlupa	60 calendar days upon receipt of the Notice to Proceed.

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____