



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-081E
PROCUREMENT OF 20 UNITS MICROSCOPE, BINOCULAR

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
>Precision of at least 2 microns	>- Precision of at least 2 2.5 microns - RETAIN
>Stage: Single specimen holder (1 pc Double specimen holder and 1 pc sheet holder)	>Stage: Single specimen holder (1 pc Double specimen holder and 1 pc sheet holder) - RETAIN

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 26, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 16th day of June, 2023 in MMCHD.

PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	MICROSCOPE, BINOCULAR	Qty./Unit	20 UNITS
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,800,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: Illumination: Built-in LED illumination system Focusing: - Coaxial coarse and fine adjustments with 200 graduations - Precision of at least 2 microns Nosepiece: Quadruple revolving nosepiece Eyepieces: Wide field 10x/20mm eyepieces Head: - 30° inclined - 360° rotatable - Interpupillary Distance: Adjustable between 55-75 mm Stage: Single specimen holder (1 pc Double specimen holder and 1 pc sheet holder) Condenser: - must be adjustable - achromatic objectives 4x, 10x, 40x and 100x - all anti-fungus treated Power: 220 VAC, 50/60 Hz Accessories: - Power Cord - LED Bulb - Dust Cover			

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
MICROSCOPE, BINOCULAR	20 UNITS	Health Facilities in National Capital Region	30 calendar days upon receipt of the Notice to Proceed

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____