

Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-082E PROCUREMENT OF 1 UNIT HEMATOLOGY ANALYZER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:		
ORIGINALTECHNICAL SPECIFICATIONS	AMENDED	
>Must be compatible with the existing reagents	>• Must be compatible with the existing	
and existing system	reagents and existing system – to	
	remove	
>Sample Volume: 20ul	• Sample Volume: 20ul — 25ul —	
	RETAIN	
>Safety standards:	Safety standards:	
- SJ/T 11363-2006 for the toxic and harmful	- SJ/T 11363-2006 for the toxic and	
substances of homogeneous materials	harmful substances of homogeneous	
	materials or EC 1272/2008 – RETAIN	
>One (1) computer set with printer	>One (1) computer set with printer	
	Processor: at least 2 cores or better	
	Ram: At least 8GB RAM	
	HardDisk: At least 256GB SSD + 1TB HDD	
	Accessories/Peripherals:	
	Keyboard, Mouse, AVR and at least 21-	
	inch 1080p Monitor with HDMI cable	
	At least Windows 11	
To remove: -Provide service machine while under		
repair		

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 26, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 16th day of June, 2023 in MMCHD.

SGD.

PRETCHELL P. TOLENTINO, MD, MCHM

Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development **TECHNICAL SPECIFICATIONS** Item No. 1 HEMATOLOGY ANALYZER Qty./Unit 1 UNIT Name of Manufacturer: Country of Origin Brand: Model: (if applicable) ABC: 1,000,000.00 PURCHASER'S SPECIFICATION STATEMENT OF COMPLIANCE TECHNICAL SPECIFICATIONS: • System: Fully Automated 5-part differential • Measurement Principle: Fluorescence staining and Flow cytometry •Measurement Parameters: 28 parameters, including 4 research parameters • Throughput: At least 60 tests/hour • Test Mode: CBC/CBC-Diff • Sample Type: Whole blood/Capillary Blood/Pre-dilution blood • Sampling Method: Manual Sampling • Sample Volume: 20ul • Sample pre-dilution: Manual injection/Automatic injection • Reagents: 4 routine reagents (Diluent, Lysing reagent, Hemolysing reagent, Fluorescent Dye) 1 cleaning solution • Quality Management - Calibrator should be FDA registered to ensure safety and quality of the machine. • Safety standards:

- SJ/T 11363-2006 for the toxic and harmful substances of

homogeneous materials

- IEC 61326 for tested electrical level

- Data Storage: At least 200,000 data
- Inclusion:
- One (1) set of start-up reagents and accessories with GHL (Glucose, Hemoglobin and Lactate) meter kit
- *One (1) computer set with printer* Processor: at least 2 cores or better

Ram: At least 8GB RAM

HardDisk: At least 256GB SSD + 1TB HDD

Accessories/Peripherals:

Keyboard, Mouse, AVR and at least 21-inch 1080p Monitor with HDMI

cable

At least Windows 11

- One (1) unit of at least 1KVA UPS

Requirements if awarded the Contract

- 1. **Completion Period**: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with **60** calendar days upon receipt of Notice to Proceed.
- 2. **Testing**: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
- 3. **Training**: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.
- 4. **Warranty**: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
- 5. **Notarized** undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
- 6. **Manuals**: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
 - a) Service manual in English language
 - b) Operation manual in English language
- 7. With "**DOH-MMCHD HFEP**" (Government Property not for sale) sticker in each unit

Source of Fund: SAA 2023-02-000687 (HFEP 2023)

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
HEMATOLOGY ANALYZER	1 UNIT	HEALTH FACILITIES IN MALABON CITY	60 calendar days upon receipt of the Notice to Proceed

Name of Company:
Address:
Signature Over Printed Name :
Telephone/Fax Number :
Email: