



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB NO. 2023-089E**  
**PROCUREMENT OF 2 UNITS AUTOMATED EXTERNAL DEFIBRILLATOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
> Energy output accuracy: $\pm 15\%$	> Energy output accuracy: $\pm 15\%$ or better
> at least five (5) pairs for adult	> at least <del>five (5)</del> ten (10) pairs for adult
> at least five (5) pairs for pedia	> at least <del>five (5)</del> ten (10) pairs for pedia

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 26, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 16<sup>th</sup> day of June, 2023 in MMCHD.

SGD  
**PRETCHELL P. TOLENTINO, MD, MCHM**  
Director III / BAC Chairperson

**Section VII. Technical Specifications**

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

**TECHNICAL SPECIFICATIONS**

Item No. <b>1</b>	<b>AUTOMATED EXTERNAL DEFIBRILATOR</b>	Qty./Unit	<b>2 Units</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC <b>300,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> <li>• Two-button Operation: On/Off, Discharge (shock button)</li> <li>• Voice Prompts: Voice Command that gives instruction to the user in operating the device</li> <li>• Automatically evaluates patient impedance for proper pad contact.</li> <li>• Automatically analyses patient condition for shockable or non-shockable rhythms</li> <li>• Patient ECG can be acquired through defibrillator pads</li> <li>• Energy output accuracy: <math>\pm 15\%</math> <b>or better</b></li> <li>• Pulse shape: Biphasic</li> <li>• Charge time: maximum of 15 seconds</li> <li>• Battery Capacity: Capable of providing at least 200 discharges at maximum energy</li> <li>• Replaceable battery</li> <li>• Visible and Audible Indicators: <ul style="list-style-type: none"> <li>- Low battery</li> <li>- Pad skin contact/disconnection</li> <li>- AED Status</li> </ul> </li> <li>• Defibrillator Pads: self-adhesive, disposable and non-polarized <ul style="list-style-type: none"> <li>- at least ten (10) pairs for adult</li> <li>- at least ten (10) pairs for pedia</li> </ul> </li> </ul>			



*Section VI. Schedule of Requirements*

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

<b>Description</b>	<b>Quantity</b>	<b>Delivery Site</b>	<b>Delivery Period</b>
<b>AUTOMATED EXTERNAL DEFIBRILLATOR</b>	2 UNITS	Health facility in National capital Region	60 calendar days upon receipt of the Notice to Proceed.

Name of Company: \_\_\_\_\_

Address: \_\_\_\_\_

Signature Over Printed Name : \_\_\_\_\_

Telephone/Fax Number : \_\_\_\_\_

Email: \_\_\_\_\_