



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB NO. 2023-093E
PROCUREMENT OF 5 UNITS PATIENT MONITOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
>Battery: - Replaceable li-ion battery - Standby duration: 10 hours - Operating duration: 5 hours	>Battery: - Standby duration: 10 hours - Operating duration: 5 hours - Duration: at least 10 hours
> VGA	> VGA or HDMI - RETAIN
>ECG > 3 lead: I, II, III > 5 lead: I, II, III, aVR, aVL, aVF, V	>ECG > 3 lead: I, II, III - REMOVE >5 lead: I, II, III, aVR, aVL, aVF, V - RETAIN
> Printer Width: at least 112mm	> Printer Width: 50mm or better at least 112mm - RETAIN

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 26, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 16th day of June, 2023 in MMCHD.

SGD
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	PATIENT MONITOR	Qty./Unit	5 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC 1,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Dimension: Manufacturer's Standard • Display: at least 12" Color LCD with LED backlight • Resolution: at least 800*600 • Battery: <ul style="list-style-type: none"> - Replaceable li-ion battery - Duration: at least 10 hours • Connectors: <ul style="list-style-type: none"> - WLAN - USB - LAN (RJ45) - VGA • Storage <ul style="list-style-type: none"> - Trend table: 168 hours - Alarm events: 100 groups - NIBP Measurements: 1600 groups - Full disclosure waveforms: 2 hours • ECG <ul style="list-style-type: none"> - 5 lead: I, II, III, aVR, aVL, aVF, V 			

<ul style="list-style-type: none"> - Accuracy: ± 1 bpm • Respiratory <ul style="list-style-type: none"> - 3 lead: RA-LA, RA-LL - 5 lead: RA-LA, RA-LL, LA-RL, LA-RL, LL-RL - Accuracy: ± 2rpm • SpO2 <ul style="list-style-type: none"> - Resolution: 1% - Accuracy: 2% • Pulse rate <ul style="list-style-type: none"> - Range: at least 20bpm - 254bpm - Accuracy: ± 3 bpm (non-motion) and ± 5 bpm (motion) or better • NIBP <ul style="list-style-type: none"> - Parameters: Systolic, Diastolic, Mean - For Adult, Pedia and Neonate • Recorder <ul style="list-style-type: none"> - Type: Thermal array - Printer Width: at least 112mm - Speed: at least 25mm/s to 50mm/s - Trace: at least 9 channel output waveform • Safety: <ul style="list-style-type: none"> - IEC 60601-1 approved - CE marking in accordance to MDD93/42/EEC • Power: 220V, 60Hz 	
<p>REQUIREMENTS, if awarded the contract</p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment. 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of 	

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
PATIENT MONITOR	5 UNITS	Health Facilities in Mandaluyong City Medical Center	60 calendar days upon receipt of the Notice to Proceed

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____