



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-096E
PROCUREMENT OF 16 UNITS OPHTHALMOSCOPE AND OTOSCOPE SET,
WALL- MOUNTED(REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 26, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 16th day of June, 2023 in MMCHD.

SGD.

PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	OPHTHALMOSCOPE AND OTOSCOPE SET, WALL-MOUNTED (REBID)	Qty./Unit	16 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC 1,120,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> ● Wall Mounted System ● Eardrum, Ear Canal and Fundus Examination ● Otoscope: 3X Magnification <ul style="list-style-type: none"> - Tunable brightness - Larger view window - Fiber optic illumination - Light: LED White or LED Sunlight - Pneumatic test function ● Direct Ophthalmoscope: Tunable Brightness <ul style="list-style-type: none"> - 5 apertures with red-free filter - 19 diopter steps: -20D ~ +20D - Light: LED White or LED Sunlight ● Power: 220V, 50/60Hz ● Modern and innovative design ● Operation time: 24 hours ● Has automatic on/off switch ● Easy wall-mounting and easy cleaning 			

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
UPHTHALMOSCOPE AND OTOSCOPE SET, WALL-MOUNTED(REBID)	16 UNITS	Health Facilities in Paranaque City	60 calendar days upon receipt of the Notice to Proceed.

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____