



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB NO. 2023-097E**  
**PROCUREMENT OF 1 UNIT INCUBATOR (FOR WATER SAMPLING) (REBID)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
> Chamber Volume: At least 340 L	> Chamber Volume: At least 300 L or better
> Shelves: At least 3 shelves provided	> Shelves: At least 2 shelves or better
> Max. load on each shelf: At least 30 kg	> Max. load on each shelf: At least 25kg or better
> Max. air exchange rate: At least 14 m <sup>3</sup> /hr	> <del>Max. air exchange rate: At least 14 m<sup>3</sup>/hr</del> - REMOVE
Add: >At least 1 or 2 chamber > Installation Area – L x W x H – 38" x 25" x 35", ± 5" each side  Laboratory Test – For microbiological testing of Water. Test includes HPC, Coliforms and Fecal Coliforms.	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on June 26, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 16<sup>th</sup> day of June, 2023 in MMCHD.

SGD  
**PRETCHELL P. TOLENTINO, MD, MCHM**  
Director III / BAC Chairperson

**Section VII. Technical Specifications**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No. 1</b>	<b>INCUBATOR (FOR WATER SAMPLING) (REBID)</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC <b>500,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS:  Chamber Volume: At least 300 L or better  At least 1 or 2 chamber  Temperature Range: Ambient + 10°C to 80°C  Temperature Stability: < ± 0.2°C  Shelves: At least 2 shelves or better  Max. load on each shelf: At least 25kg or better  Power Supply: 220V, 60Hz  Installation Area – L x W x H – 38" x 25" x 35", ± 5" each side  Laboratory Test – For microbiological testing of Water. Test includes HPC, Coliforms and Fecal Coliforms.			
Requirements if awarded the Contract  <b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <b>60</b> calendar days upon receipt of Notice to Proceed.  <b>2. Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.  <b>3. Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.  <b>4. Warranty:</b> Warranty certificate for two (2) years on parts and service.			



**Section VI. Schedule of Requirements**

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

<b>Description</b>	<b>Quantity</b>	<b>Delivery Site</b>	<b>Delivery Period</b>
<b>INCUBATOR (FOR WATER SAMPLING) (REBID)</b>	1 UNIT	Health facilities in Marikina City	60 calendar days upon receipt of the Notice to Proceed

Name of Company: \_\_\_\_\_

Address: \_\_\_\_\_

Signature Over Printed Name : \_\_\_\_\_

Telephone/Fax Number : \_\_\_\_\_

Email: \_\_\_\_\_