



SUPPLEMENTAL/ BID BULLETIN NO. 1

B2024 – 077E

PROCUREMENT OF 5 UNITS RADIANT WARMER (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Ceramic heater elements and golden reflectors.	Ceramic heater elements and golden reflectors or its equivalent	Ceramic heater elements and golden reflectors or its equivalent GRANTED
With at least 10 heater level.	With 5 heater level or better	With 5 heater level or better
With at least 2 heater elements located at the Radiant warmer	With at least 1 heater elements located at the Radiant warmer or better	With at least 1 heater elements located at the Radiant warmer or better
With 30 watts day and 9 watts night examination light	With 30 watts day and 9 watts night examination light	With 30 watts day and 9 watts night examination light
With temperature measuring range of at least 15 degrees to 42 degrees	With temperature measuring range of at least 15 degrees to 42 degrees or Manufacturer Standard	With temperature measuring range of at least 15 degrees to 42 degrees or Manufacturer Standard
Variable Height Adjustment (VHA): At least 290mm	Variable Height Adjustment (VHA): At least 290mm or Manufacturer Standard	Variable Height Adjustment (VHA): At least 290mm or Manufacturer Standard
Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed	Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed	Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed


Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 22nd day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 13th day of August 2024 in MMCHD

Approved by:



JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	RADIANT WARMER	Qty./Unit	5 UNITS
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,250,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Movable Lamp mechanism that automatically focuses radiant warmth on the baby. • Lamp can be positioned on the overhead left or right side of the equipment. • <i>Ceramic heater elements and golden reflectors or its equivalent.</i> • Radiant warmth with Two (2) brightness level Examination lamp features. • <i>With 5 heater level or better</i> • Radiant Warmer can be positioned from left to right direction for up to 90 degrees. • <i>With at least 1 heater elements located at the Radiant warmer or better</i> • <i>With examination light</i> • Radiant Power at Heat Level 3 between the bed and radiant warmer: 10mW/cm² • Radiant Power at Heat Level 10 between the bed and radiant warmer: 30mW/cm² • Lamps (Work Light): 30 Watts • Lamps (Night Light): 9 Watts • Minimum clearance between top edge of Radiant warmer and Ceiling: >50cm • Temperature <ul style="list-style-type: none"> - Capable of measuring baby's central or peripheral temperature. - Manual mode - Skin servo mode - <i>With temperature measuring range of at least 15 degrees to 42 degrees or Manufacturer Standard</i> - Accuracy (Sensor): ±0.1 degrees Celsius 			-

<ul style="list-style-type: none"> - Setting Range: 30 degrees Celsius to 38.5 degrees Celsius • <i>Variable Height Adjustment (VHA): At least 290mm or Manufacturer Standard</i> • Standard Features <ul style="list-style-type: none"> - Swivel cupboards located at right or left side of the machine. - Manual bed-tilt feature Head up position is 20 degrees. - Manual bed-tilt feature Head down position is 15 degrees. - The inner side walls must be 7cm. - Equipped with Standard X-ray tray. • Accessories <ul style="list-style-type: none"> - One (1) unit Skin temperature probe - One (1) unit Power cable - One (1) unit Power adapter 	
<p>Requirements if awarded the Contract</p> <p>1 Completion Period: <i>The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed</i></p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. Warranty</p> <p>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</p> <p>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</p> <p>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the</p>	<p>1..</p>

equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "**DOH-MMCHD HFEP**"(Government Property not for sale) sticker in each unit.

X-X

Source of Fund: SAA 2023-02-000687 (HFEP 2023 ConAp)
Recipient: Caloocan City Medical Center