



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB2024 – 080E**

**PROCUREMENT OF VARIOUS HOSPITAL BEDS (LOT BIDDING) (REBID)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
<b>Item NO. 1: Examination Table</b>		
No changes as stipulated in the technical specification		
<b>Item No. 2: Transfer Stretcher</b>		
Dimensions: L x W x H = 2000mm x 700mm x 700mm, ± 50mm on each dimension	Dimensions: L x W x H = 2000mm x 700mm x 700mm, ± 100mm on each dimension - for clarification with end user	Dimensions: L x W x H = 2000mm x 700mm x 700mm, ± 100mm on each dimension - GRANTED
<b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <b>30</b> calendar days upon receipt of Notice to Proceed	. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.	<b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.
<b>Item No. 3: Bed, Mechanical, 3Crank with Mattress</b>		
Sliding full side rails	Sliding 3/4 side rails - for clarification with end user	Sliding 3/4 side rails - for clarification with end user
Caster w/ 4" diameter	Caster w/ 4" diameter or Manufacturer Standard	Caster w/ 4" diameter or Manufacturer Standard
<b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be	<b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to	<b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to

completed within <b>30</b> calendar days upon receipt of Notice to Proceed	Proceed.	Proceed.
<b>Item No 4: Wheeled Detachable Stretcher</b>		
No changes as stipulated in the technical specification		

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 22nd day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 13<sup>th</sup> day of August 2024 in MMCHD

Approved by:



**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson

Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	<b>Examination Table</b>	Qty./Unit	<b>15 UNITS</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>750,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> <li>• Material: Powdered Coated Metal/Steel or better</li> <li>• With detachable stirrups</li> <li>• Has PVC or washable vinyl mattress</li> <li>• With tray or drain pan</li> <li>• With table that has at least 2 drawers and 2 cabinets</li> <li>• Drawer &amp; cabinet dimensions: Manufacturer's Standard</li> <li>• Dimensions (L x W x H): at least 58" x 20" x 34"</li> <li>• Accessories: <ul style="list-style-type: none"> <li>- Foot stool</li> <li>- Assembly tools</li> </ul> </li> </ul>			
<p><b>Requirements if awarded the Contract</b></p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Warranty:</b> Warranty certificate for one (1) year on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. <b>Notarized undertaking</b> that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> <li>a) Service manual in English language</li> <li>b) Operation manual in English language</li> </ul> <p>7. With "<b>DOH-MMCHD HFEP</b>"(Government Property not for sale) sticker</p>			







<ul style="list-style-type: none"> <li>• Has a separable stretcher that can be removed and be used as a separate stretcher</li> <li>• Mattress: At least 3 cm thick, waterproof seamless mattress</li> <li>• Has 3 belts</li> <li>• Castor wheels: At least 100 mm diameter castors</li> <li>• Max Load: at least 160kg</li> <li>• Dimensions: L x W x H = 2000mm x 600mm x 850mm, ± 50mm in every dimension, adjustable height must be at least in range or bigger</li> <li>• Adjustable height: 300mm to 850mm</li> </ul>	
<p><b>Requirements</b> if awarded the Contract</p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> <li>a) Service manual in English language</li> <li>b) Operation manual in English language</li> </ul> <p>7. With "<b>DOH-MMCHD HFEP</b>" (Government Property not for sale) sticker in each unit</p> <p>x-x</p> <p>Source of Fund: SAA 2024-02-001002 (HFEP 2024)</p> <p>Recipient: Health Facilities in Marikina City</p>	