



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 030E

#### PROCUREMENT OF 8 UNITS PULSE OXIMETER WITH ADULT AND PEDIA PROBE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>From</b>	<b>To</b>
Power Supply: 220V, 50/60Hz or Battery operated	Power Supply: Dual Power Operation (220VAC or Autovolt, 50Hz/60Hz and Battery should be rechargeable)
Safety Standard: EN60601-1	Safety Standard: EN60601-1 or its equivalent
Compact side configuration	Detachable Handle
Display: LED	Display: LCD or better


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 11<sup>th</sup> day of January 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 4<sup>th</sup> day of January, 2022 in MMCHD

Approved by:

  
ALELI ANNIE GRACE P. SUDIACAL, MD, MPH  
Director III / BAC Chairperson 8